



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01384	Calib. Date: 12/11/2017
Certification File No.:	01385	Calib. No.: 00031
Linearity File No.:	01386	Cert. Date: 12/11/2017
Solution File No.:	01411	Cert. No.: 00026
Sequential File No.:	01411	Lin. Date: 12/11/2017
		Lin. No.: 00026
		Soln. Date: 04/15/2018
		Soln. No.: 00196
		File Date: 04/15/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 0665

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:17D		
Control 1 EC	0.100%	19:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:18D		
Control 2 EC	0.100%	19:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:19D		
Control 3 EC	0.099%	19:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRIFFIN

First Name: RUSSELL

MI: -

Signature: _____

Badge No.: 96

Date: 04/15/2018