



Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARSA-0031 |
| Location: | STAFFORD TOWNSHIP P.D. | | |
| Calibration File No.: | 01384 | Calib. Date: | 12/11/2017 |
| Certification File No.: | 01385 | Cert. Date: | 12/11/2017 |
| Linearity File No.: | 01386 | Lin. Date: | 12/11/2017 |
| Solution File No.: | 01391 | Soln. Date: | 01/11/2018 |
| Sequential File No.: | 01391 | File Date: | 01/11/2018 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUD S3-0015 |
| Control Solution %: | 0.100% | | | Expires: | 06/28/2019 |
| Solution Control Lot: | 17200 | | | Bottle No.: | 0324 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 20:55S | | |
| Control 1 EC | 0.103% | 20:56S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 20:56S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:56S | | |
| Control 2 EC | 0.102% | 20:57S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 20:57S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:58S | | |
| Control 3 EC | 0.101% | 20:58S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 20:58S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:59S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SMITH

First Name: CHRISTOPHER

MI: M

Signature: _____

Badge No.: 102

Date: 01/11/2018