



Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01419	Calib. Date: 05/31/2018
Certification File No.:	01420	Calib. No.: 00032
Linearity File No.:	01421	Cert. Date: 05/31/2018
Solution File No.:	01434	Cert. No.: 00027
Sequential File No.:	01434	Lin. Date: 05/31/2018
		Lin. No.: 00027
		Soln. Date: 07/22/2018
		Soln. No.: 00200
		File Date: 07/22/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0014
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 0663

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:25D		
Control 1 EC	0.100%	18:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26D		
Control 2 EC	0.100%	18:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:27D		
Control 3 EC	0.100%	18:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SMITHMAN

First Name: ERIK

MI: T

Signature: _____

Badge No.: 103

Date: 07/22/2018