



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01419	Calib. Date:	05/31/2018
Certification File No.:	01420	Cert. Date:	05/31/2018
Linearity File No.:	01421	Lin. Date:	05/31/2018
Solution File No.:	01427	Soln. Date:	06/21/2018
Sequential File No.:	01427	File Date:	06/21/2018

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUD S3-0014
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	0325

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:56D		
Control 1 EC	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 2 EC	0.099%	10:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 3 EC	0.099%	10:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: EIKER

First Name: KARL

MI: D

Signature: _____

P. Eiker #113

Badge No.: 113

Date: 06/21/2018