



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01384	Calib. Date:	12/11/2017
Certification File No.:	01385	Cert. No.:	00031
Linearity File No.:	01386	Cert. Date:	12/11/2017
Solution File No.:	01401	Lin. No.:	00026
Sequential File No.:	01401	Lin. Date:	12/11/2017
		Soln. Date:	03/14/2018
		Soln. No.:	00195
		File Date:	03/14/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0015
Solution Control Lot:	17200	Expires:	06/28/2019
		Bottle No.:	0313

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:17S		
Control 1 EC	0.102%	23:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:18S		
Control 2 EC	0.101%	23:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:19S		
Control 3 EC	0.100%	23:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____

Badge No.: 111

Date: 03/14/2018