



Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARSA-0031 |
| Location: | STAFFORD TOWNSHIP P.D. | |
| Calibration File No.: | 01384 | Calib. Date: 12/11/2017 |
| Certification File No.: | 01385 | Calib. No.: 00031 |
| Linearity File No.: | 01386 | Cert. Date: 12/11/2017 |
| Solution File No.: | 01416 | Cert. No.: 00026 |
| Sequential File No.: | 01416 | Lin. Date: 12/11/2017 |
| | | Lin. No.: 00026 |
| | | Soln. Date: 05/27/2018 |
| | | Soln. No.: 00197 |
| | | File Date: 05/27/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUD S3-0015 |
| Solution Control Lot: | 17200 | Expires: 06/28/2019 |
| | | Bottle No.: 0566 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:04D | | |
| Control 1 EC | 0.100% | 14:05D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 14:05D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:05D | | |
| Control 2 EC | 0.100% | 14:06D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 14:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:06D | | |
| Control 3 EC | 0.100% | 14:07D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 14:07D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:08D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: _____

Badge No.: 98

Date: 05/27/2018