

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01419 Calib. Date: 05/31/2018 Calib. No.: 00032
Certification File No.: 01420 Cert. Date: 05/31/2018 Cert. No.: 00027
Linearity File No.: 01421 Lin. Date: 05/31/2018 Lin. No.: 00027
Solution File No.: 01438 Soln. Date: 08/24/2018 Soln. No.: 00201
Sequential File No.: 01438 File Date: 08/24/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0171

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:47D		
Control 1 EC	0.099%	01:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:49D		
Control 2 EC	0.098%	01:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:50D		
Control 3 EC	0.098%	01:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DEMARCO

First Name: GIOVANNI

MI: -

Signature: *PD Demarco #104*

Badge No.: 104

Date: 08/24/2018