

✓

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARSA-0031  
Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01419 Calib. Date: 05/31/2018 Calib. No.: 00032  
Certification File No.: 01420 Cert. Date: 05/31/2018 Cert. No.: 00027  
Linearity File No.: 01421 Lin. Date: 05/31/2018 Lin. No.: 00027  
Solution File No.: 01442 Soln. Date: 09/24/2018 Soln. No.: 00202  
Sequential File No.: 01442 File Date: 09/24/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014  
Control Solution %: 0.100% Expires: 05/10/2020  
Solution Control Lot: 18150 Bottle No.: 0816

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:22D		
Control 1 EC	0.099%	03:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:23D		
Control 2 EC	0.098%	03:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:24D		
Control 3 EC	0.098%	03:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: POTTER

First Name: MARC

MI: D

Signature: \_\_\_\_\_

 A 105

Badge No.: 105

Date: 09/24/2018