

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01448	Calib. Date:	11/19/2018
Certification File No.:	01449	Cert. Date:	11/19/2018
Linearity File No.:	01450	Lin. Date:	11/19/2018
Solution File No.:	01451	Soln. Date:	11/19/2018
Sequential File No.:	01451	File Date:	11/19/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0015
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0767

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:35S		
Control 1 EC	0.100%	10:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		
Control 2 EC	0.099%	10:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:37S		
Control 3 EC	0.099%	10:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXK P2-304 (DN)

Changed By:

Last Name: NAPOLITANO	First Name: DAVID	MI: M.
Signature: <u><i>For II Del M. Napolitano</i></u>	Badge No.: 7237	Date: 11/19/2018
	7237	