

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01448 Calib. Date: 11/19/2018 Calib. No.: 00033
Certification File No.: 01449 Cert. Date: 11/19/2018 Cert. No.: 00028
Linearity File No.: 01450 Lin. Date: 11/19/2018 Lin. No.: 00028
Solution File No.: 01455 Soln. Date: 12/17/2018 Soln. No.: 00205
Sequential File No.: 01455 File Date: 12/17/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0632

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:50S		
Control 1 EC	0.099%	17:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51S		
Control 2 EC	0.098%	17:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		
Control 3 EC	0.098%	17:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SMITHMAN

First Name: ERIK

MI: T

Signature: _____



Badge No.: 103

Date: 12/17/2018