



Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01448	Calib. Date: 11/19/2018
Certification File No.:	01449	Calib. No.: 00033
Linearity File No.:	01450	Cert. Date: 11/19/2018
Solution File No.:	01460	Cert. No.: 00028
Sequential File No.:	01460	Lin. Date: 11/19/2018
		Lin. No.: 00028
		Soln. Date: 01/17/2019
		Soln. No.: 00206
		File Date: 01/17/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0873

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:13S		
Control 1 EC	0.100%	07:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14S		
Control 2 EC	0.098%	07:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:15S		
Control 3 EC	0.099%	07:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CONNER

First Name: JACOB

MI: N

Signature: _____

Badge No.: 109

Date: 01/17/2019