

✓

# Calibrating Unit New Standard Solution Report

|                         |                        |              |            |             |              |
|-------------------------|------------------------|--------------|------------|-------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARSA-0031  |             |              |
| Location:               | STAFFORD TOWNSHIP P.D. |              |            |             |              |
| Calibration File No.:   | 01448                  | Calib. Date: | 11/19/2018 | Calib. No.: | 00033        |
| Certification File No.: | 01449                  | Cert. Date:  | 11/19/2018 | Cert. No.:  | 00028        |
| Linearity File No.:     | 01450                  | Lin. Date:   | 11/19/2018 | Lin. No.:   | 00028        |
| Solution File No.:      | 01465                  | Soln. Date:  | 02/03/2019 | Soln. No.:  | 00207        |
| Sequential File No.:    | 01465                  | File Date:   | 02/03/2019 |             |              |
| <br>                    |                        |              |            |             |              |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34      | Serial No.: | DDUD S3-0015 |
| Control Solution %:     | 0.100%                 |              |            | Expires:    | 05/10/2020   |
| Solution Control Lot:   | 18150                  |              |            | Bottle No.: | 0815         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 11:09S |                |                     |
| Control 1 EC      | 0.100% | 11:10S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.098% | 11:10S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:11S |                |                     |
| Control 2 EC      | 0.098% | 11:11S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 11:11S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:12S |                |                     |
| Control 3 EC      | 0.098% | 11:13S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 11:13S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:13S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: \_\_\_\_\_



Badge No.: 98

Date: 02/03/2019