

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C

Serial No.: ARSA-0031

Location:

STAFFORD TOWNSHIP P.D.

Calibration File No.: 01448

Calib. Date: 11/19/2018

Calib. No.: 00033

Certification File No.: 01449

Cert. Date: 11/19/2018

Cert. No.: 00028

Linearity File No.: 01450

Lin. Date: 11/19/2018

Lin. No.: 00028

Solution File No.: 01477

Soln. Date: 03/30/2019

Soln. No.: 00210

Sequential File No.: 01477

File Date: 03/30/2019

Calibrating Unit: WET

Model No.: CU-34

Serial No.: DDUD S3-0015

Control Solution %: 0.100%

Expires: 05/10/2020

Solution Control Lot: 18150

Bottle No.: 0163

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:28S		
Control 1 EC	0.100%	21:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		
Control 2 EC	0.098%	21:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31S		
Control 3 EC	0.098%	21:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: 

Badge No.: 112

Date: 03/30/2019