

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01448	Calib. Date:	11/19/2018
Certification File No.:	01449	Cert. Date:	11/19/2018
Linearity File No.:	01450	Lin. Date:	11/19/2018
Solution File No.:	01483	Soln. Date:	05/01/2019
Sequential File No.:	01483	File Date:	05/01/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUD S3-0015
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	0347

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:42D		
Control 1 EC	0.099%	03:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:44D		
Control 2 EC	0.098%	03:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:45D		
Control 3 EC	0.097%	03:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: POTTER

First Name: MARC

MI: D

Signature: \_\_\_\_\_

Badge No.: 105

Date: 05/01/2019