

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01448	Calib. No.: 00033
Certification File No.:	01449	Cert. No.: 00028
Linearity File No.:	01450	Lin. No.: 00028
Solution File No.:	01491	Soln. No.: 00212
Sequential File No.:	01491	File Date: 05/11/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0694

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:43D		
Control 1 EC	0.099%	00:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44D		
Control 2 EC	0.098%	00:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45D		
Control 3 EC	0.098%	00:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____



Badge No.: 111

Date: 05/11/2019