

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01494	Calib. Date: 05/15/2019
Certification File No.:	01495	Calib. No.: 00034
Linearity File No.:	01496	Cert. Date: 05/15/2019
Solution File No.:	01503	Cert. No.: 00029
Sequential File No.:	01503	Lin. Date: 05/15/2019
		Lin. No.: 00029
		Soln. Date: 05/31/2019
		Soln. No.: 00214
		File Date: 05/31/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0463

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:29D		
Control 1 EC	0.099%	01:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:30D		
Control 2 EC	0.098%	01:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:31D		
Control 3 EC	0.098%	01:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NICCOLI

First Name: JOSEPH

MI: R

Signature: _____



116

Badge No.: 116

Date: 05/31/2019