

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01494 Calib. Date: 05/15/2019 Calib. No.: 00034
Certification File No.: 01495 Cert. Date: 05/15/2019 Cert. No.: 00029
Linearity File No.: 01496 Lin. Date: 05/15/2019 Lin. No.: 00029
Solution File No.: 01521 Soln. Date: 08/06/2019 Soln. No.: 00220
Sequential File No.: 01521 File Date: 08/06/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0695

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:22D		
Control 1 EC	0.098%	00:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23D		
Control 2 EC	0.097%	00:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		
Control 3 EC	0.098%	00:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SMITH

First Name: CHRISTOPHER

MI: M

Signature: _____



Badge No.: 102

Date: 08/06/2019