

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01494 Calib. Date: 05/15/2019 Calib. No.: 00034
Certification File No.: 01495 Cert. Date: 05/15/2019 Cert. No.: 00029
Linearity File No.: 01496 Lin. Date: 05/15/2019 Lin. No.: 00029
Solution File No.: 01530 Soln. Date: 08/24/2019 Soln. No.: 00221
Sequential File No.: 01530 File Date: 08/24/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0874

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:11D		
Control 1 EC	0.099%	02:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:12D		
Control 2 EC	0.097%	02:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14D		
Control 3 EC	0.099%	02:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 112

Date: 08/24/2019