



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01536	Calib. Date: 09/13/2019
Certification File No.:	01537	Calib. No.: 00035
Linearity File No.:	01538	Cert. Date: 09/13/2019
Solution File No.:	01551	Cert. No.: 00030
Sequential File No.:	01551	Lin. Date: 09/13/2019
		Lin. No.: 00030
		Soln. Date: 10/31/2019
		Soln. No.: 00224
		File Date: 10/31/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 1198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:30S		
Control 1 EC	0.097%	17:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:31S		
Control 2 EC	0.097%	17:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:32S		
Control 3 EC	0.098%	17:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____

Badge No.: 111

Date: 10/31/2019