

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01536	Calib. Date:	09/13/2019
Certification File No.:	01537	Cert. Date:	09/13/2019
Linearity File No.:	01538	Lin. Date:	09/13/2019
Solution File No.:	01557	Soln. Date:	12/04/2019
Sequential File No.:	01557	File Date:	12/04/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0015
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1034

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:59S		
Control 1 EC	0.100%	05:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:00S		
Control 2 EC	0.098%	06:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01S		
Control 3 EC	0.098%	06:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POTTER

First Name: MARC

MI: D

Signature:



Badge No.: 105

Date: 12/04/2019