

# Calibrating Unit

## New Standard Solution Report



<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01536	Calib. Date:	09/13/2019
Certification File No.:	01537	Cert. No.:	00035
Linearity File No.:	01538	Cert. Date:	09/13/2019
Solution File No.:	01566	Lin. No.:	00030
Sequential File No.:	01566	Soln. Date:	12/21/2019
		Soln. No.:	00226
		File Date:	12/21/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0015
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1087

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:06S		
Control 1 EC	0.097%	01:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:07S		
Control 2 EC	0.097%	01:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:09S		
Control 3 EC	0.099%	01:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NICCOLI

First Name: JOSEPH

MI: R

Signature: \_\_\_\_\_

Badge No.: 116

Date: 12/21/2019