

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARSA-0031  
Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01536 Calib. Date: 09/13/2019 Calib. No.: 00035  
Certification File No.: 01537 Cert. Date: 09/13/2019 Cert. No.: 00030  
Linearity File No.: 01538 Lin. Date: 09/13/2019 Lin. No.: 00030  
Solution File No.: 01572 Soln. Date: 02/07/2020 Soln. No.: 00228  
Sequential File No.: 01572 File Date: 02/07/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1014

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:24S		
Control 1 EC	0.097%	08:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		
Control 2 EC	0.096%	08:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27S		
Control 3 EC	0.097%	08:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: EIKER

First Name: KARL

MI: D

Signature: \_\_\_\_\_

*Karl Eiker #113*

Badge No.: 113

Date: 02/07/2020