

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARSA-0031  
Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01579 Calib. Date: 06/10/2020 Calib. No.: 00038  
Certification File No.: 01580 Cert. Date: 06/10/2020 Cert. No.: 00031  
Linearity File No.: 01581 Lin. Date: 06/10/2020 Lin. No.: 00031  
Solution File No.: 01587 Soln. Date: 07/21/2020 Soln. No.: 00232  
Sequential File No.: 01587 File Date: 07/21/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1028

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:34D		
Control 1 EC	0.099%	07:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35D		
Control 2 EC	0.098%	07:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:36D		
Control 3 EC	0.099%	07:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: POTTER

First Name: MARC

MI: D

Signature: 

Badge No.: 105

Date: 07/21/2020