



Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
 Location: STAFFORD TOWNSHIP P.D.
 Calibration File No.: 01609 Calib. Date: 11/05/2020 Calib. No.: 00039
 Certification File No.: 01610 Cert. Date: 11/05/2020 Cert. No.: 00032
 Linearity File No.: 01611 Lin. Date: 11/05/2020 Lin. No.: 00032
 Solution File No.: 01618 Soln. Date: 12/04/2020 Soln. No.: 00237
 Sequential File No.: 01618 File Date: 12/04/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 1032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:17S		
Control 1 EC	0.100%	05:18S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:18S		
Control 2 EC	0.100%	05:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:19S		
Control 3 EC	0.100%	05:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____

Badge No.: 111

Date: 12/04/2020