



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	Calib. No.:	00038
Calibration File No.:	01579	Cert. Date:	06/10/2020
Certification File No.:	01580	Lin. Date:	06/10/2020
Linearity File No.:	01581	Soln. Date:	10/28/2020
Solution File No.:	01608	File Date:	10/28/2020
Sequential File No.:	01608		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0014
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1080

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:50S		
Control 1 EC	0.099%	22:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:51S		
Control 2 EC	0.098%	22:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52S		
Control 3 EC	0.099%	22:53S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

Badge No.: 112

Date: 10/28/2020