



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01579	Calib. Date: 06/10/2020
Certification File No.:	01580	Calib. No.: 00038
Linearity File No.:	01581	Cert. Date: 06/10/2020
Solution File No.:	01601	Cert. No.: 00031
Sequential File No.:	01601	Lin. Date: 06/10/2020
		Lin. No.: 00031
		Soln. Date: 09/28/2020
		Soln. No.: 00234
		File Date: 09/28/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0014
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1072

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:59D		
Control 1 EC	0.099%	16:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00D		
Control 2 EC	0.098%	16:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01D		
Control 3 EC	0.098%	16:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STANZIANO

First Name: JOSEPH

MI: T

Signature: \_\_\_\_\_

Badge No.: 114

Date: 09/28/2020