

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	Calib. No.:	00039
Calibration File No.:	01609	Calib. Date:	11/05/2020
Certification File No.:	01610	Cert. Date:	11/05/2020
Linearity File No.:	01611	Lin. Date:	11/05/2020
Solution File No.:	01631	Soln. Date:	02/02/2021
Sequential File No.:	01631	File Date:	02/02/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0014
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1077

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:55S		
Control 1 EC	0.099%	11:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56S		
Control 2 EC	0.099%	11:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57S		
Control 3 EC	0.099%	11:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: \_\_\_\_\_



Badge No.: 98

Date: 02/02/2021