



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01609	Calib. Date: 11/05/2020
Certification File No.:	01610	Calib. No.: 00039
Linearity File No.:	01611	Cert. Date: 11/05/2020
Solution File No.:	01623	Cert. No.: 00032
Sequential File No.:	01623	Lin. Date: 11/05/2020
		Lin. No.: 00032
		Soln. Date: 12/31/2020
		Soln. No.: 00238
		File Date: 12/31/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0014
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1031

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:24S		
Control 1 EC	0.100%	23:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25S		
Control 2 EC	0.099%	23:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26S		
Control 3 EC	0.099%	23:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____

Badge No.: 111

Date: 12/31/2020