

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARSA-0031  
Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01609 Calib. Date: 11/05/2020 Calib. No.: 00039  
Certification File No.: 01610 Cert. Date: 11/05/2020 Cert. No.: 00032  
Linearity File No.: 01611 Lin. Date: 11/05/2020 Lin. No.: 00032  
Solution File No.: 01638 Soln. Date: 03/27/2021 Soln. No.: 00241  
Sequential File No.: 01638 File Date: 03/27/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015  
Control Solution %: 0.100% Expires: 09/09/2021  
Solution Control Lot: 19240 Bottle No.: 1071

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:43S		
Control 1 EC	0.100%	00:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45S		
Control 2 EC	0.099%	00:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46S		
Control 3 EC	0.099%	00:46S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NICCOLI

First Name: JOSEPH

MI: R

Signature: 

Badge No.: 116

Date: 03/27/2021