

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01644 Calib. Date: 04/21/2021 Calib. No.: 00040
Certification File No.: 01645 Cert. Date: 04/21/2021 Cert. No.: 00033
Linearity File No.: 01646 Lin. Date: 04/21/2021 Lin. No.: 00033
Solution File No.: 01653 Soln. Date: 05/20/2021 Soln. No.: 00243
Sequential File No.: 01653 File Date: 05/20/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0914

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:19D		
Control 1 EC	0.099%	20:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:20D		
Control 2 EC	0.099%	20:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:22D		
Control 3 EC	0.099%	20:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 112

Date: 05/20/2021