

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01644	Calib. Date:	04/21/2021
Certification File No.:	01645	Cert. Date:	04/21/2021
Linearity File No.:	01646	Lin. Date:	04/21/2021
Solution File No.:	01666	Soln. Date:	06/14/2021
Sequential File No.:	01666	File Date:	06/14/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUD S3-0015
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	1076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:19D		
Control 1 EC	0.099%	00:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:20D		
Control 2 EC	0.098%	00:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:21D		
Control 3 EC	0.098%	00:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SMITH

First Name: CHRISTOPHER

MI: M

Signature: _____



Badge No.: 102

Date: 06/14/2021