

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARSA-0031 |
| Location: | STAFFORD TOWNSHIP P.D. | |
| Calibration File No.: | 01644 | Calib. Date: 04/21/2021 |
| Certification File No.: | 01645 | Calib. No.: 00040 |
| Linearity File No.: | 01646 | Cert. Date: 04/21/2021 |
| Solution File No.: | 01675 | Cert. No.: 00033 |
| Sequential File No.: | 01675 | Lin. Date: 04/21/2021 |
| | | Lin. No.: 00033 |
| | | Soln. Date: 07/08/2021 |
| | | Soln. No.: 00245 |
| | | File Date: 07/08/2021 |

| | | | |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDUD S3-0015 |
| Control Solution %: | 0.100% | | Expires: 09/09/2021 |
| Solution Control Lot: | 19240 | | Bottle No.: 1164 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 19:03D | | |
| Control 1 EC | 0.098% | 19:03D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 19:03D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:04D | | |
| Control 2 EC | 0.098% | 19:04D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 19:04D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:05D | | |
| Control 3 EC | 0.098% | 19:06D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 19:06D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 19:06D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 112

Date: 07/08/2021