

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01644	Calib. Date: 04/21/2021
Certification File No.:	01645	Calib. No.: 00040
Linearity File No.:	01646	Cert. Date: 04/21/2021
Solution File No.:	01694	Cert. No.: 00033
Sequential File No.:	01694	Lin. Date: 04/21/2021
		Lin. No.: 00033
		Soln. Date: 08/30/2021
		Soln. No.: 00247
		File Date: 08/30/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	19380	Expires: 12/16/2021
		Bottle No.: 0225

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:48D		
Control 1 EC	0.097%	04:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:49D		
Control 2 EC	0.097%	04:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:50D		
Control 3 EC	0.097%	04:51D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRIFFIN

First Name: RUSSELL

MI: -

Signature: _____



Badge No.: 96

Date: 08/30/2021