

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARSA-0031  
Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01701 Calib. Date: 10/06/2021 Calib. No.: 00041  
Certification File No.: 01702 Cert. Date: 10/06/2021 Cert. No.: 00034  
Linearity File No.: 01703 Lin. Date: 10/06/2021 Lin. No.: 00034  
Solution File No.: 01711 Soln. Date: 12/11/2021 Soln. No.: 00250  
Sequential File No.: 01711 File Date: 12/11/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 0897

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:21S		
Control 1 EC	0.102%	01:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22S		
Control 2 EC	0.100%	01:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23S		
Control 3 EC	0.101%	01:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NICCOLI

First Name: JOSEPH

MI: R

Signature: 

Badge No.: 116

Date: 12/11/2021