

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01701	Calib. Date: 10/06/2021
Certification File No.:	01702	Calib. No.: 00041
Linearity File No.:	01703	Cert. Date: 10/06/2021
Solution File No.:	01707	Cert. No.: 00034
Sequential File No.:	01707	Lin. Date: 10/06/2021
		Lin. No.: 00034
		Soln. Date: 11/01/2021
		Soln. No.: 00249
		File Date: 11/01/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0014
Solution Control Lot:	19380	Expires: 12/16/2021
		Bottle No.: 1297

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:41S		
Control 1 EC	0.100%	12:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42S		
Control 2 EC	0.099%	12:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43S		
Control 3 EC	0.100%	12:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: EIKER

First Name: KARL

MI: D

Signature: _____

Karl Eiker #113

Badge No.: 113

Date: 11/01/2021