

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARSA-0031 |
| Location: | STAFFORD TOWNSHIP P.D. | |
| Calibration File No.: | 01701 | Calib. Date: 10/06/2021 |
| Certification File No.: | 01702 | Calib. No.: 00041 |
| Linearity File No.: | 01703 | Cert. Date: 10/06/2021 |
| Solution File No.: | 01716 | Cert. No.: 00034 |
| Sequential File No.: | 01716 | Lin. Date: 10/06/2021 |
| | | Lin. No.: 00034 |
| | | Soln. Date: 01/15/2022 |
| | | Soln. No.: 00251 |
| | | File Date: 01/15/2022 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUD S3-0014 |
| Solution Control Lot: | 21360 | Expires: 09/01/2023 |
| | | Bottle No.: 1039 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 03:26S | | |
| Control 1 EC | 0.100% | 03:27S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 03:27S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:28S | | |
| Control 2 EC | 0.099% | 03:28S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 03:28S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:29S | | |
| Control 3 EC | 0.100% | 03:29S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 03:29S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 03:30S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: _____



Badge No.: 98

Date: 01/15/2022