

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01701 Calib. Date: 10/06/2021 Calib. No.: 00041
Certification File No.: 01702 Cert. Date: 10/06/2021 Cert. No.: 00034
Linearity File No.: 01703 Lin. Date: 10/06/2021 Lin. No.: 00034
Solution File No.: 01729 Soln. Date: 03/11/2022 Soln. No.: 00253
Sequential File No.: 01729 File Date: 03/11/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1043

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:41S		
Control 1 EC	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:42S		
Control 2 EC	0.100%	15:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44S		
Control 3 EC	0.100%	15:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKEY

First Name: MARC

MI: R

Signature: _____



Badge No.: 111

Date: 03/11/2022