

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARSA-0031 |
| Location: | STAFFORD TOWNSHIP P.D. | | |
| Calibration File No.: | 01701 | Calib. Date: | 10/06/2021 |
| Certification File No.: | 01702 | Cert. Date: | 10/06/2021 |
| Linearity File No.: | 01703 | Lin. Date: | 10/06/2021 |
| Solution File No.: | 01723 | Soln. Date: | 02/15/2022 |
| Sequential File No.: | 01723 | File Date: | 02/15/2022 |
| Calibrating Unit: WET | | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUD S3-0014 |
| Solution Control Lot: | 21360 | Expires: | 09/01/2023 |
| | | Bottle No.: | 1357 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 01:45S | | |
| Control 1 EC | 0.100% | 01:46S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 01:46S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:46S | | |
| Control 2 EC | 0.099% | 01:47S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 01:47S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:48S | | |
| Control 3 EC | 0.099% | 01:48S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 01:48S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:49S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____

Badge No.: 112

Date: 02/15/2022