

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01731	Calib. Date:	03/15/2022
Certification File No.:	01732	Cert. Date:	03/15/2022
Linearity File No.:	01733	Lin. Date:	03/15/2022
Solution File No.:	01741	Soln. Date:	04/08/2022
Sequential File No.:	01741	File Date:	04/08/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0015
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1260

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:43D		
Control 1 EC	0.098%	22:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44D		
Control 2 EC	0.097%	22:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		
Control 3 EC	0.097%	22:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: _____



Badge No.: 98

Date: 04/08/2022