

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01731	Calib. Date: 03/15/2022
Certification File No.:	01732	Calib. No.: 00042
Linearity File No.:	01733	Cert. Date: 03/15/2022
Solution File No.:	01750	Cert. No.: 00035
Sequential File No.:	01750	Lin. Date: 03/15/2022
		Lin. No.: 00035
		Soln. Date: 05/04/2022
		Soln. No.: 00258
		File Date: 05/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 0995

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:35D		
Control 1 EC	0.098%	00:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36D		
Control 2 EC	0.097%	00:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:37D		
Control 3 EC	0.097%	00:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 112

Date: 05/04/2022