

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01731 Calib. Date: 03/15/2022 Calib. No.: 00042
Certification File No.: 01732 Cert. Date: 03/15/2022 Cert. No.: 00035
Linearity File No.: 01733 Lin. Date: 03/15/2022 Lin. No.: 00035
Solution File No.: 01748 Soln. Date: 04/28/2022 Soln. No.: 00257
Sequential File No.: 01748 File Date: 04/28/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1358

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:50D		
Control 1 EC	0.098%	00:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:51D		
Control 2 EC	0.097%	00:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:53D		
Control 3 EC	0.097%	00:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POTTER

First Name: MARC

MI: D

Signature: 

Badge No.: 105

Date: 04/28/2022