

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01731 Calib. Date: 03/15/2022 Calib. No.: 00042
Certification File No.: 01732 Cert. Date: 03/15/2022 Cert. No.: 00035
Linearity File No.: 01733 Lin. Date: 03/15/2022 Lin. No.: 00035
Solution File No.: 01759 Soln. Date: 07/11/2022 Soln. No.: 00261
Sequential File No.: 01759 File Date: 07/11/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 0887

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:02D		
Control 1 EC	0.098%	13:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		
Control 2 EC	0.097%	13:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04D		
Control 3 EC	0.097%	13:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BUSH

First Name: JACKMAN

MI: R

Signature: _____



Badge No.: 127

Date: 07/11/2022