

# Calibrating Unit

## New Standard Solution Report

### Equipment

Alcotest 7110 MKIII-C  
STAFFORD TOWNSHIP P.D.  
Serial No.: ARSA-0031

Location: STAFFORD TOWNSHIP P.D.  
Calibration File No.: 01769 Calib. Date: 08/19/2022 Calib. No.: 00043  
Certification File No.: 01770 Cert. Date: 08/19/2022 Cert. No.: 00036  
Linearity File No.: 01771 Lin. Date: 08/19/2022 Lin. No.: 00036  
Solution File No.: 01781 Soln. Date: 09/16/2022 Soln. No.: 00264  
Sequential File No.: 01781 File Date: 09/16/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 1089

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 01:30D |                |                     |
| Control 1 EC      | 0.100% | 01:31D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 01:31D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:31D |                |                     |
| Control 2 EC      | 0.099% | 01:32D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 01:32D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:33D |                |                     |
| Control 3 EC      | 0.099% | 01:33D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 01:33D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:34D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KUNDER

First Name: EDWARD

MI: L

Signature: \_\_\_\_\_



Badge No.: 98

Date: 09/16/2022