

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01769 Calib. Date: 08/19/2022 Calib. No.: 00043
Certification File No.: 01770 Cert. Date: 08/19/2022 Cert. No.: 00036
Linearity File No.: 01771 Lin. Date: 08/19/2022 Lin. No.: 00036
Solution File No.: 01786 Soln. Date: 10/17/2022 Soln. No.: 00265
Sequential File No.: 01786 File Date: 10/17/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0015
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 0849

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:35D		
Control 1 EC	0.099%	02:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:36D		
Control 2 EC	0.099%	02:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:37D		
Control 3 EC	0.099%	02:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: BRYAN

MI: S

Signature: 

Badge No.: 126

Date: 10/17/2022