

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01769	Calib. Date: 08/19/2022
Certification File No.:	01770	Calib. No.: 00043
Linearity File No.:	01771	Cert. Date: 08/19/2022
Solution File No.:	01792	Cert. No.: 00036
Sequential File No.:	01792	Lin. Date: 08/19/2022
		Lin. No.: 00036
		Soln. Date: 11/13/2022
		Soln. No.: 00266
		File Date: 11/13/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1469

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:25S		
Control 1 EC	0.100%	21:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		
Control 2 EC	0.099%	21:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:28S		
Control 3 EC	0.099%	21:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NICCOLI

First Name: JOSEPH

MI: R

Signature: _____



#116

Badge No.: 116

Date: 11/13/2022