

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01731	Calib. Date: 03/15/2022
Certification File No.:	01732	Calib. No.: 00042
Linearity File No.:	01733	Cert. Date: 03/15/2022
Solution File No.:	01764	Cert. No.: 00035
Sequential File No.:	01764	Lin. Date: 03/15/2022
		Lin. No.: 00035
		Soln. Date: 08/06/2022
		Soln. No.: 00262
		File Date: 08/06/2022

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUD S3-0015
Control Solution %:	0.100%		Expires: 09/01/2023
Solution Control Lot:	21360		Bottle No.: 1413

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:55D		
Control 1 EC	0.097%	08:55D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		
Control 2 EC	0.097%	08:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		
Control 3 EC	0.097%	08:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: \_\_\_\_\_



Badge No.: 111

Date: 08/06/2022