

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.		
Calibration File No.:	01814	Calib. Date:	01/24/2023
Certification File No.:	01815	Cert. No.:	00044
Linearity File No.:	01816	Cert. Date:	01/24/2023
Solution File No.:	01824	Lin. Date:	01/24/2023
Sequential File No.:	01824	Lin. No.:	00037
		Soln. Date:	03/26/2023
		Soln. No.:	00271
		File Date:	03/26/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUD S3-0014
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	0274

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:21S		
Control 1 EC	0.100%	10:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		
Control 2 EC	0.099%	10:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23S		
Control 3 EC	0.099%	10:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: EIKER

First Name: KARL

MI: D

Signature: _____

Karl Eiker #113

Badge No.: 113

Date: 03/26/2023