

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01814	Calib. Date: 01/24/2023
Certification File No.:	01815	Calib. No.: 00044
Linearity File No.:	01816	Cert. Date: 01/24/2023
Solution File No.:	01829	Cert. No.: 00037
Sequential File No.:	01829	Lin. Date: 01/24/2023
		Lin. No.: 00037
		Soln. Date: 04/25/2023
		Soln. No.: 00272
		File Date: 04/25/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0014
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0333

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:03D		
Control 1 EC	0.100%	23:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:04D		
Control 2 EC	0.100%	23:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06D		
Control 3 EC	0.100%	23:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SMITH

First Name: CHRISTOPHER

MI: T

Signature: _____



Badge No.: 120

Date: 04/25/2023