

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01814 Calib. Date: 01/24/2023 Calib. No.: 00044
Certification File No.: 01815 Cert. Date: 01/24/2023 Cert. No.: 00037
Linearity File No.: 01816 Lin. Date: 01/24/2023 Lin. No.: 00037
Solution File No.: 01833 Soln. Date: 05/21/2023 Soln. No.: 00273
Sequential File No.: 01833 File Date: 05/21/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 0126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:06D		
Control 1 EC	0.100%	02:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:07D		
Control 2 EC	0.099%	02:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08D		
Control 3 EC	0.099%	02:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 112

Date: 05/21/2023