



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01769	Calib. Date: 08/19/2022
Certification File No.:	01770	Calib. No.: 00043
Linearity File No.:	01771	Cert. Date: 08/19/2022
Solution File No.:	01809	Cert. No.: 00036
Sequential File No.:	01809	Lin. Date: 08/19/2022
		Lin. No.: 00036
		Soln. Date: 01/13/2023
		Soln. No.: 00268
		File Date: 01/13/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1042

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:36S		
Control 1 EC	0.100%	02:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:38S		
Control 2 EC	0.099%	02:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:39S		
Control 3 EC	0.099%	02:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

Badge No.: 112

Date: 01/13/2023